



Cascade Engineering Services, Inc.



ONSITE DAILY SUMMARY LIST

ONSITE DATE: Wednesday, March 15, 2023

ELLENSBURG PD

CURRENT PO:

Asset / ID Number	MANUFACTURER	MODEL NUMBER	DESCRIPTION	DEPARTMENT	RESU
FF12385	KUSTOM SIGNALS, INC.	FALCON	SMD	N/A	PAS:
FF12396	KUSTOM SIGNALS, INC.	FALCON	SMD	N/A	PAS:
PYT546004242	MPH	PYTHON III	SMD	N/A	PAS:
PYT546004243	MPH	PYTHON II	SMD	N/A	PAS:
PYT546004244	MPH	PYTHON III	SMD	N/A	PAS:
PYT546005120	MPH	PYTHON II	SMD	N/A	PAS:
PYT846003563	MPH	PYTHON III	SMD	N/A	PAS:
PYT846006198	MPH	PYTHON III	SMD	N/A	PAS:

Number Processed: 8

Customer Signature: _____

Date: _____



Metrology Laboratory
 Cascade Engineering Services, Inc.

VOICE: 425-895-8617
 FAX: 425-895-8620
 6640 185th Ave NE
 Redmond, WA 98052

SHIPPER/RECEIVER

COMPANY ELLENSBURG PD DATE SHIPPED 3/15/2023
 ADDRESS 100 N. PEARL ST.
 CITY ELLENSBURG
 STATE WA ZIP 98926
 CALIBRATION PURCHASE ORDER _____ REPAIR PURCHASE ORDE _____

DHL FEDEX UPS USPS CES
 OTHER: _____

DEPARTMENT	ID NUMBER	MANUFACTURER	MODEL	SERIAL NUMBER	CALIBRATION INTERVAL	SERVICE REQUIRED		
						CAL	REPAIR	DATA
1	10236	MPH	880010	10236	12	X		
2	LD079874	APPLIED CONCEPTS	STALKER	LD079874	12	X		
3								
4								
5								
6								
7								
8								
9								
10								
11								
12								
13								
14								
15								
16								

ACCESSORIES, SPECIAL INSTRUCTIONS, COMMENTS

CLIENT AUTHORIZED SIGNATURE _____ RECEIVED BY J. RUSSAL GRAY IV



Cascade Engineering Services, Inc.



6640 185th Ave NE, Redmond, WA 98052

T.425.895.8617, F.425.702.9358

ON-SITE CERTIFICATE #: OS3-00330291

CERTIFICATE OF CALIBRATION

STANDARD CALIBRATION

ELLENSBURG POLICE DEPARTMENT

100 N. PEARL STREET ELLENSBURG, WA 98926

This certifies that the instrument listed herein was calibrated by Cascade Engineering Services' Calibration Laboratory, which is fully accredited in accordance with the recognized International Standards ISO/IEC 17025:2017 General Requirements for the Competence of Testing and Calibration Laboratories. Cascade Engineering Services' Calibration Laboratory also meets the requirements of ANSI/NC SL Z540-1-1994 and any additional program requirements in the field of calibration. Standards used to perform this calibration are certified by or traceable to NIST, natural physical constants, consensus standards or derived by the ratio type of calibrations. All calibrations are performed to manufacturer's specifications unless otherwise noted. Standard Calibration, while still traceable, does not meet all requirements for an Accredited Calibration per ISO/IEC 17025:2017, that is As Found data for equipment in tolerance and Measurement Uncertainties are not recorded. This certificate shall not be reproduced, except in full, without prior written approval of the laboratory.

DESCRIPTION:

ASSET #: PYT546004244
MANUFACTURER: MPH
DEPARTMENT: N/A
ENVIRONMENT: 72.0 °F/48.1 %RH
CAL INTERVAL: 12 MONTHS

SMD

SERIAL NUMBER: PYT546004244
MODEL NUMBER: PYTHON III
LOCATION: 11-P3
BASIC ACCURACY: REFERENCE MFRS SPECIFICATIONS
DUE DATE: March 15, 2024

EQUIPMENT CONDITION AS RECEIVED

Initial testing found this equipment to be "IN TOLERANCE", as defined by the basic accuracy stated above.

EQUIPMENT CONDITION AS DELIVERED

At the completion of the calibration, measured values were "IN TOLERANCE", as defined by the basic accuracy stated above.

TUNING FORK(S) SUPPLIED WITH THIS DEVICE

DESCRIPTION	SERIAL NUMBER	RATED SPEED	FREQUENCY
TUNING FORK ONE	290101	35 M.P.H.	2524 Hz
TUNING FORK TWO	290554	65 M.P.H.	4690 Hz

Antenna 1 SN: PYT315012004, Frequency: 24.159 GHz

STANDARD(S) USED FOR CERTIFICATION

I.D.	MODEL	MANUFACTURER	DESCRIPTION	DUE DATE
MET1387	VOCAR HR	DB INNOVATIONS	HAND HELD RADAR CERTIFICATION SYSTEM	12/15/2023
MET1388	VOCAR HR WAND	DB INNOVATIONS	VOCAR HR WAND	12/15/2023

PROCEDURE(S) USED FOR CERTIFICATION

DOCUMENT ID	DESCRIPTION	REV	REV DATE
CP-SMD-001	RADAR AND LIDAR CALIBRATION PROCEDURE	NEW	11/14/2019

CERTIFICATION NOTES

I certify (or declare) under penalty of perjury under the laws of the State of Washington that the above information is true and correct

PERFORMED BY

METROLOGIST JOHN GRAY

LOCATION: ELLENSBURG, WA

CALIBRATION DATE: Wednesday, March 15, 2023



Cascade Engineering Services, Inc.



6640 185th Ave NE, Redmond, WA 98052

T.425.895.8617, F.425.702.9358

ON-SITE CERTIFICATE #: OS3-00330290

CERTIFICATE OF CALIBRATION

STANDARD CALIBRATION

ELLENSBURG POLICE DEPARTMENT

100 N. PEARL STREET ELLENSBURG, WA 98926

This certifies that the instrument listed herein was calibrated by Cascade Engineering Services' Calibration Laboratory, which is fully accredited in accordance with the recognized International Standards ISO/IEC 17025:2017 General Requirements for the Competence of Testing and Calibration Laboratories. Cascade Engineering Services' Calibration Laboratory also meets the requirements of ANSI/NC SL Z540-1-1994 and any additional program requirements in the field of calibration. Standards used to perform this calibration are certified by or traceable to NIST, natural physical constants, consensus standards or derived by the ratio type of calibrations. All calibrations are performed to manufacturer's specifications unless otherwise noted. Standard Calibration, while still traceable, does not meet all requirements for an Accredited Calibration per ISO/IEC 17025:2017, that is As Found data for equipment in tolerance and Measurement Uncertainties are not recorded. This certificate shall not be reproduced, except in full, without prior written approval of the laboratory.

DESCRIPTION:

ASSET #: PYT546004243
MANUFACTURER: MPH
DEPARTMENT: N/A
ENVIRONMENT: 72.0 °F/48.1 %RH
CAL INTERVAL: 12 MONTHS

SMD

SERIAL NUMBER: PYT546004243
MODEL NUMBER: PYTHON II
LOCATION: 15-P7
BASIC ACCURACY: REFERENCE MFRS SPECIFICATIONS
DUE DATE: March 15, 2024

EQUIPMENT CONDITION AS RECEIVED

Initial testing found this equipment to be "IN TOLERANCE", as defined by the basic accuracy stated above.

EQUIPMENT CONDITION AS DELIVERED

At the completion of the calibration, measured values were "IN TOLERANCE", as defined by the basic accuracy stated above.

TUNING FORK(S) SUPPLIED WITH THIS DEVICE

DESCRIPTION	SERIAL NUMBER	RATED SPEED	FREQUENCY
TUNING FORK ONE	307534	35 M.P.H.	2534 Hz
TUNING FORK TWO	307560	65 M.P.H.	4702 Hz

Antenna 1 SN: PYT315012003, Frequency: 24.152 GHz

Antenna 2 SN: PYT315018529, Frequency: 24.204 GHz

STANDARD(S) USED FOR CERTIFICATION

I.D.	MODEL	MANUFACTURER	DESCRIPTION	DUE DATE
MET1387	VOCAR HR	DB INNOVATIONS	HAND HELD RADAR CERTIFICATION SYSTEM	12/15/2023
MET1388	VOCAR HR WAND	DB INNOVATIONS	VOCAR HR WAND	12/15/2023

PROCEDURE(S) USED FOR CERTIFICATION

DOCUMENT ID	DESCRIPTION	REV	REV DATE
CP-SMD-001	RADAR AND LIDAR CALIBRATION PROCEDURE	NEW	11/14/2019

CERTIFICATION NOTES

I certify (or declare) under penalty of perjury under the laws of the State of Washington that the above information is true and correct

PERFORMED BY 
 METROLOGIST JOHN GRAY

LOCATION: ELLENSBURG, WA
 CALIBRATION DATE: Wednesday, March 15, 2023



Cascade Engineering Services, Inc.



6640 185th Ave NE, Redmond, WA 98052

T.425.895.8617, F.425.702.9358

ON-SITE CERTIFICATE #: OS3-00330289

CERTIFICATE OF CALIBRATION

STANDARD CALIBRATION

ELLENSBURG POLICE DEPARTMENT

100 N. PEARL STREET ELLENSBURG, WA 98926

This certifies that the instrument listed herein was calibrated by Cascade Engineering Services' Calibration Laboratory, which is fully accredited in accordance with the recognized International Standards ISO/IEC 17025:2017 General Requirements for the Competence of Testing and Calibration Laboratories. Cascade Engineering Services' Calibration Laboratory also meets the requirements of ANSI/NC SL Z540-1-1994 and any additional program requirements in the field of calibration. Standards used to perform this calibration are certified by or traceable to NIST, natural physical constants, consensus standards or derived by the ratio type of calibrations. All calibrations are performed to manufacturer's specifications unless otherwise noted. Standard Calibration, while still traceable, does not meet all requirements for an Accredited Calibration per ISO/IEC 17025:2017, that is As Found data for equipment in tolerance and Measurement Uncertainties are not recorded. This certificate shall not be reproduced, except in full, without prior written approval of the laboratory.

DESCRIPTION:

SMD

ASSET #:	PYT546004242	SERIAL NUMBER:	PYT546004242
MANUFACTURER:	MPH	MODEL NUMBER:	PYTHON III
DEPARTMENT:	N/A	LOCATION:	TRAFFIC
ENVIRONMENT:	72.0 °F/48.1 %RH	BASIC ACCURACY:	REFERENCE MFRS SPECIFICATIONS
CAL INTERVAL:	12 MONTHS	DUE DATE:	March 15, 2024

EQUIPMENT CONDITION AS RECEIVED

Initial testing found this equipment to be "IN TOLERANCE", as defined by the basic accuracy stated above.

EQUIPMENT CONDITION AS DELIVERED

At the completion of the calibration, measured values were "IN TOLERANCE", as defined by the basic accuracy stated above.

TUNNING FORK(S) SUPPLIED WITH THIS DEVICE

DESCRIPTION	SERIAL NUMBER	RATED SPEED	FREQUENCY
TUNING FORK ONE	978326	35 M.P.H.	2538 Hz
TUNING FORK TWO	978335	65 M.P.H.	4712 Hz

Antenna 1 SN: PYT315012002, Frequency: 24.162 GHz

STANDARD(S) USED FOR CERTIFICATION

I.D.	MODEL	MANUFACTURER	DESCRIPTION	DUE DATE
MET1387	VOCAR HR	DB INNOVATIONS	HAND HELD RADAR CERTIFICATION SYSTEM	12/15/2023
MET1388	VOCAR HR WAND	DB INNOVATIONS	VOCAR HR WAND	12/15/2023

PROCEDURE(S) USED FOR CERTIFICATION

DOCUMENT ID	DESCRIPTION	REV	REV DATE
CP-SMD-001	RADAR AND LIDAR CALIBRATION PROCEDURE	NEW	11/14/2019

CERTIFICATION NOTES

I certify (or declare) under penalty of perjury under the laws of the State of Washington that the above information is true and correct

PERFORMED BY

METROLOGIST JOHN GRAY

LOCATION: ELLENSBURG, WA

CALIBRATION DATE: Wednesday, March 15, 2023



Cascade Engineering Services, Inc.



6640 185th Ave NE, Redmond, WA 98052

T.425.895.8617, F.425.702.9358

ON-SITE CERTIFICATE #: OS3-00330288

CERTIFICATE OF CALIBRATION

STANDARD CALIBRATION

ELLENSBURG POLICE DEPARTMENT

100 N. PEARL STREET ELLENSBURG, WA 98926

This certifies that the instrument listed herein was calibrated by Cascade Engineering Services' Calibration Laboratory, which is fully accredited in accordance with the recognized International Standards ISO/IEC 17025:2017 General Requirements for the Competence of Testing and Calibration Laboratories. Cascade Engineering Services' Calibration Laboratory also meets the requirements of ANSI/NC SL Z540-1-1994 and any additional program requirements in the field of calibration. Standards used to perform this calibration are certified by or traceable to NIST, natural physical constants, consensus standards or derived by the ratio type of calibrations. All calibrations are performed to manufacturer's specifications unless otherwise noted. Standard Calibration, while still traceable, does not meet all requirements for an Accredited Calibration per ISO/IEC 17025:2017, that is As Found data for equipment in tolerance and Measurement Uncertainties are not recorded. This certificate shall not be reproduced, except in full, without prior written approval of the laboratory.

DESCRIPTION:

SMD

ASSET #: PYT846006198
 MANUFACTURER: MPH
 DEPARTMENT: N/A
 ENVIRONMENT: 72.0 °F/48.1 %RH
 CAL INTERVAL: 12 MONTHS

SERIAL NUMBER: PYT846006198
 MODEL NUMBER: PYTHON III
 LOCATION: 17-P8
 BASIC ACCURACY: REFERENCE MFRS SPECIFICATIONS
 DUE DATE: March 15, 2024

EQUIPMENT CONDITION AS RECEIVED

Initial testing found this equipment to be "IN TOLERANCE", as defined by the basic accuracy stated above.

EQUIPMENT CONDITION AS DELIVERED

At the completion of the calibration, measured values were "IN TOLERANCE", as defined by the basic accuracy stated above.

TUNNING FORK(S) SUPPLIED WITH THIS DEVICE

DESCRIPTION	SERIAL NUMBER	RATED SPEED	FREQUENCY
TUNING FORK ONE	2302	35 M.P.H.	2532 Hz
TUNING FORK TWO	2196	65 M.P.H.	4708 Hz

Antenna 1 SN: PYT315013697, Frequency: 24.158 GHz

STANDARD(S) USED FOR CERTIFICATION

I.D.	MODEL	MANUFACTURER	DESCRIPTION	DUE DATE
MET1387	VOCAR HR	DB INNOVATIONS	HAND HELD RADAR CERTIFICATION SYSTEM	12/15/2023
MET1388	VOCAR HR WAND	DB INNOVATIONS	VOCAR HR WAND	12/15/2023

PROCEDURE(S) USED FOR CERTIFICATION

DOCUMENT ID	DESCRIPTION	REV	REV DATE
CP-SMD-001	RADAR AND LIDAR CALIBRATION PROCEDURE	NEW	11/14/2019

CERTIFICATION NOTES

I certify (or declare) under penalty of perjury under the laws of the State of Washington that the above information is true and correct

PERFORMED BY 
 METROLOGIST JOHN GRAY

LOCATION: ELLENSBURG, WA
 CALIBRATION DATE: Wednesday, March 15, 2023



Cascade Engineering Services, Inc.



6640 185th Ave NE, Redmond, WA 98052

T.425.895.8617, F.425.702.9358

ON-SITE CERTIFICATE #: OS3-00330287

CERTIFICATE OF CALIBRATION

STANDARD CALIBRATION

ELLENSBURG POLICE DEPARTMENT

100 N. PEARL STREET ELLENSBURG, WA 98926

This certifies that the instrument listed herein was calibrated by Cascade Engineering Services' Calibration Laboratory, which is fully accredited in accordance with the recognized International Standards ISO/IEC 17025:2017 General Requirements for the Competence of Testing and Calibration Laboratories. Cascade Engineering Services' Calibration Laboratory also meets the requirements of ANSI/NCSL Z540-1-1994 and any additional program requirements in the field of calibration. Standards used to perform this calibration are certified by or traceable to NIST, natural physical constants, consensus standards or derived by the ratio type of calibrations. All calibrations are performed to manufacturer's specifications unless otherwise noted. Standard Calibration, while still traceable, does not meet all requirements for an Accredited Calibration per ISO/IEC 17025:2017, that is As Found data for equipment in tolerance and Measurement Uncertainties are not recorded. This certificate shall not be reproduced, except in full, without prior written approval of the laboratory.

DESCRIPTION:

ASSET #: PYT546005120
MANUFACTURER: MPH
DEPARTMENT: N/A
ENVIRONMENT: 72.0 °F/48.1 %RH
CAL INTERVAL: 12 MONTHS

SMD

SERIAL NUMBER: PYT546005120
MODEL NUMBER: PYTHON II
LOCATION: 16-P4
BASIC ACCURACY: REFERENCE MFRS SPECIFICATIONS
DUE DATE: March 15, 2024

EQUIPMENT CONDITION AS RECEIVED

Initial testing found this equipment to be "IN TOLERANCE", as defined by the basic accuracy stated above.

EQUIPMENT CONDITION AS DELIVERED

At the completion of the calibration, measured values were "IN TOLERANCE", as defined by the basic accuracy stated above.

TUNING FORK(S) SUPPLIED WITH THIS DEVICE

DESCRIPTION	SERIAL NUMBER	RATED SPEED	FREQUENCY
TUNING FORK ONE	857240	35 M.P.H.	2530 Hz
TUNING FORK TWO	071109	65 M.P.H.	4696 Hz

Antenna 1 SN: PYT315013698, Frequency: 24.15 GHz

Antenna 2 SN: PYT315018527, Frequency: 24.185 GHz

STANDARD(S) USED FOR CERTIFICATION

I.D.	MODEL	MANUFACTURER	DESCRIPTION	DUE DATE
MET1387	VOCAR HR	DB INNOVATIONS	HAND HELD RADAR CERTIFICATION SYSTEM	12/15/2023
MET1388	VOCAR HR WAND	DB INNOVATIONS	VOCAR HR WAND	12/15/2023

PROCEDURE(S) USED FOR CERTIFICATION

DOCUMENT ID	DESCRIPTION	REV	REV DATE
CP-SMD-001	RADAR AND LIDAR CALIBRATION PROCEDURE	NEW	11/14/2019

CERTIFICATION NOTES

I certify (or declare) under penalty of perjury under the laws of the State of Washington that the above information is true and correct

PERFORMED BY

METROLOGIST JOHN GRAY

LOCATION: ELLENSBURG, WA

CALIBRATION DATE: Wednesday, March 15, 2023



Cascade Engineering Services, Inc.



6640 185th Ave NE, Redmond, WA 98052
T.425.895.8617, F.425.702.9358

ON-SITE CERTIFICATE #: OS3-00330286

CERTIFICATE OF CALIBRATION

STANDARD CALIBRATION

ELLENSBURG POLICE DEPARTMENT

100 N. PEARL STREET ELLENSBURG, WA 98926

This certifies that the instrument listed herein was calibrated by Cascade Engineering Services' Calibration Laboratory, which is fully accredited in accordance with the recognized International Standards ISO/IEC 17025:2017 General Requirements for the Competence of Testing and Calibration Laboratories. Cascade Engineering Services' Calibration Laboratory also meets the requirements of ANSI/NCCL Z540-1-1994 and any additional program requirements in the field of calibration. Standards used to perform this calibration are certified by or traceable to NIST, natural physical constants, consensus standards or derived by the ratio type of calibrations. All calibrations are performed to manufacturer's specifications unless otherwise noted. Standard Calibration, while still traceable, does not meet all requirements for an Accredited Calibration per ISO/IEC 17025:2017, that is As Found data for equipment in tolerance and Measurement Uncertainties are not recorded. This certificate shall not be reproduced, except in full, without prior written approval of the laboratory.

DESCRIPTION:

ASSET #: FF12396
MANUFACTURER: KUSTOM SIGNALS, INC.
DEPARTMENT: N/A
ENVIRONMENT: 72.0 °F/48.1 %RH
CAL INTERVAL: 12 MONTHS

SMD

SERIAL NUMBER: FF12396
MODEL NUMBER: FALCON
LOCATION: N/A
BASIC ACCURACY: REFERENCE MFRS SPECIFICATIONS
DUE DATE: March 15, 2024

EQUIPMENT CONDITION AS RECEIVED

Initial testing found this equipment to be "IN TOLERANCE", as defined by the basic accuracy stated above.

EQUIPMENT CONDITION AS DELIVERED

At the completion of the calibration, measured values were "IN TOLERANCE", as defined by the basic accuracy stated above.

TUNING FORK(S) SUPPLIED WITH THIS DEVICE

DESCRIPTION	SERIAL NUMBER	RATED SPEED	FREQUENCY
TUNING FORK ONE	24639	50 M.P.H.	3670 Hz
TUNING FORK TWO	N/A		

Antenna 1 SN: HANDHELD, Frequency: 24.171 GHz

STANDARD(S) USED FOR CERTIFICATION

I.D.	MODEL	MANUFACTURER	DESCRIPTION	DUE DATE
MET1387	VOCAR HR	DB INNOVATIONS	HAND HELD RADAR CERTIFICATION SYSTEM	12/15/2023
MET1388	VOCAR HR WAND	DB INNOVATIONS	VOCAR HR WAND	12/15/2023

PROCEDURE(S) USED FOR CERTIFICATION


DOCUMENT ID	DESCRIPTION	REV	REV DATE
CP-SMD-001	RADAR AND LIDAR CALIBRATION PROCEDURE	NEW	11/14/2019

CERTIFICATION NOTES

I certify (or declare) under penalty of perjury under the laws of the State of Washington that the above information is true and correct

PERFORMED BY 
METROLOGIST- JOHN GRAY

LOCATION: ELLENSBURG, WA
CALIBRATION DATE: Wednesday, March 15, 2023



CERTIFICATE OF CALIBRATION

STANDARD CALIBRATION

ELLENSBURG POLICE DEPARTMENT

100 N. PEARL STREET ELLENSBURG, WA 98926

This certifies that the instrument listed herein was calibrated by Cascade Engineering Services' Calibration Laboratory, which is fully accredited in accordance with the recognized International Standards ISO/IEC 17025:2017 General Requirements for the Competence of Testing and Calibration Laboratories. Cascade Engineering Services' Calibration Laboratory also meets the requirements of ANSI/NCSL Z540-1-1994 and any additional program requirements in the field of calibration. Standards used to perform this calibration are certified by or traceable to NIST, natural physical constants, consensus standards or derived by the ratio type of calibrations. All calibrations are performed to manufacturer's specifications unless otherwise noted. Standard Calibration, while still traceable, does not meet all requirements for an Accredited Calibration per ISO/IEC 17025:2017, that is As Found data for equipment in tolerance and Measurement Uncertainties are not recorded. This certificate shall not be reproduced, except in full, without prior written approval of the laboratory.

DESCRIPTION:

ASSET #: PYT846003563

MANUFACTURER: MPH

DEPARTMENT: N/A

ENVIRONMENT: 72.0 °F/48.1 %RH

CAL INTERVAL: 12 MONTHS

SMD

SERIAL NUMBER: PYT846003563

MODEL NUMBER: PYTHON III

LOCATION: 18-P6

BASIC ACCURACY: REFERENCE MFRS SPECIFICATIONS

DUE DATE: March 15, 2024

EQUIPMENT CONDITION AS RECEIVED

Initial testing found this equipment to be "IN TOLERANCE", as defined by the basic accuracy stated above.

EQUIPMENT CONDITION AS DELIVERED

At the completion of the calibration, measured values were "IN TOLERANCE", as defined by the basic accuracy stated above.

TUNNING FORK(S) SUPPLIED WITH THIS DEVICE

DESCRIPTION	SERIAL NUMBER	RATED SPEED	FREQUENCY
TUNING FORK ONE	077371	35 M.P.H.	2530 Hz
TUNING FORK TWO	077383	65 M.P.H.	4698 Hz

Antenna 1 SN: PYT855004681, Frequency: 24.181 GHz

STANDARD(S) USED FOR CERTIFICATION

I.D.	MODEL	MANUFACTURER	DESCRIPTION	DUE DATE
MET1387	VOCAR HR	DB INNOVATIONS	HAND HELD RADAR CERTIFICATION SYSTEM	12/15/2023
MET1388	VOCAR HR WAND	DB INNOVATIONS	VOCAR HR WAND	12/15/2023

PROCEDURE(S) USED FOR CERTIFICATION

DOCUMENT ID	DESCRIPTION	REV	REV DATE
CP-SMD-001	RADAR AND LIDAR CALIBRATION PROCEDURE	NEW	11/14/2019

CERTIFICATION NOTES

I certify (or declare) under penalty of perjury under the laws of the State of Washington that the above information is true and correct

PERFORMED BY 
 METROLOGIST JOHN GRAY

LOCATION: ELLENSBURG, WA
CALIBRATION DATE: Wednesday, March 15, 2023